


(D)	PLOT SCALE: 1=1	DWG. SCALE: 1	
-----	-----------------	---------------	--



# NOTES:

1. LEAD PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED SHEET. FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL BE LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION-QQ1201. CABINET SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: NBS HANDBOOK 111 AND 114, ANS N43.2 AND DOE ORDER 5480.11.
2. FINISH - PAINT EXTERIOR SURFACES, EXCEPT AS NOTED, WITH PRATT & LAMBERT #2127 (SNO-FLAKE WHITE, SEMI-GLOSS, EPOXY BASE) OR EQUIVALENT. SURFACE PREPARATION PER A.N.L. SPECIFICATION #4105-000 24.
3. DIMENSIONS IN BRACKETS [ ] ARE MILLIMETERS AND ARE FOR REFERENCE ONLY.
4. SOME HIDDEN LINES MAY BE OMITTED FOR CLARITY.

ITEM	DWG/PART NUMBER	NOMENCLATURE OR DESCRIPTION	MATERIAL / SPEC	QTY
<div> <div> UNLESS OTHERWISE SPECIFIED: ALL DIMENSIONS ARE IN INCHES TOLERANCES: DECIMALS .005 FRACTIONS 1/16 SURFACE ROUGHNESS 125 </div> <div> LOG NUMBER <b>A2240600</b> DRAWN BY <b>M. PATAKY</b> DATE <b>4/5/95</b> DESIGNED BY <b>SANCHEZ/PATAKY</b> DATE <b>4/5/95</b> DESIGNER <b>SANCHEZ/PATAKY</b> DATE <b>4/5/95</b> RESPONSIBLE ENGINEER <b>SANCHEZ/PATAKY</b> DATE <b>4/5/95</b> APPROVED/RELEASED <b>SANCHEZ/PATAKY</b> DATE <b>4/5/95</b> </div> <div> THIS DRAWING IS THE PROPERTY OF <b>ARGONNE NATIONAL LABORATORY</b> TITLE <b>ADVANCED PHOTON SOURCE U6 ID SHIELDED CABINET DIAGNOSTICS ASSEMBLY ROOF LABYRINTH</b> SCALE 1:1 SHEET 1 of 1 DRAWING NUMBER <b>P4105090606-900010-00</b> </div> </div>				
<div> <div> SEE NOTES </div> <div> DO NOT SCALE DRAWING </div> </div>				